

Certificate

Certificate no.

CU 72407511 0001

License Holder:

Heidolph Scientific Products GmbH
Walpersdorfer Str. 12
91126 Schwabach
Deutschland

Manufacturing Plant:

DE22ZTJM

Report Number:

DE22ZTJM 004

Client Reference:

Walter Lohmann

Certification acc. to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Product Information

Certified Product:

Vacuum Pumps for Laboratory Use

Model Designation:

Hei-VAC Valve Control; (1)
Hei-VAC Valve Industrial (2)

Technical Data:

Input : (1) 230V 50/60Hz 1.8A; or
ratings (1) 100-115V 50/60Hz 3.4A / 120V 60Hz 3.4A;
(2) 100-115V 50/60Hz 5.7A /
200-230V 50/60Hz 3.0A
Protection: Class I; IP 40/Type 1(UL50E)

Remarks:

License holder name changed. Certificate replaces CU72228652 01

Appendix:

-/-



Date of issue:

2024-10-25
(yr/mo/day)

TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

 TÜVRheinland®