

Certificate

Certificate no.

CU 72407521 0001

License Holder:

Heidolph Scientific Products GmbH
Walpersdorfer Str. 12
91126 Schwabach
Deutschland

Manufacturing Plant:

32195323

Report Number:

32195323 009

Client Reference:

Walter Lohmann

Certification acc. to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Product Information

Certified Product:

Diaphragm vacuum pump for laboratory use

Model Designation:

Hei-VAC Valve Tec

Technical Data:

Input voltage: AC 120V; 60Hz /
AC 100-115V; 50/60Hz
Input current: 1.6-1.7A
Protection: class I; IP40/Type 1 (UL 50E)

Remarks:

License holder name changed. Certificate replaces CU72211805 01

Appendix:



Date of issue:

2024-10-25
(yr/mo/day)

TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

 TÜVRheinland®